

## EC Declaration

Document no. DZ. PC. PR.28.Z2.ENG

Manufacturer: **APLISENS S.A.**,  
**ul. Morelowa 7, 03-192 Warszawa**

We hereby acknowledge, that:

Pressure transmitters **PCE-28, PCE-28P**, differential pressure transmitters **PRE-28** in all versions comply with directive

**EMC – 2004/108/WE**

The following standard was applied PN-EN 61326-1:2006

Pressure transmitters model **PCE-28**  
in **PED** version comply with directive

• **PED – 97/23/WE**

Transmitters in PED version are marked with the following certificate marking :

**CE-PED-B-APL 001-09-POL** - re. **PCE-28**, module **B + D** (category IV)

Certificates were issued by Notified Body no. 0062, Bureau Veritas ; 67/71 Bd du Château, 92200 Neuilly-Sur-Seine, France

Pressure transmitters models **PCE-28, PCE-28P**, differential pressure transmitters **PRE-28**  
in Intrinsically Safe version comply with directive

• **ATEX – 94/9/WE**

Transmitters in Intrinsically Safe version are marked with the following certificate marking:

II 1/2G Ex ia IIC T4/T5/T6

I M1, Ex ia I

Certificate no. **KDB 08ATEX002X**

Certificate was issued by Notified Body no. 1453, Central Mining Institute, Plac Gwarków 1, 40-166 Katowice.

The following standards were applied: PN-EN60079-0:2006, PN-EN60079-11:2007, PN-EN60079-26:2005, PN-EN50303:2004.

• **Marine version**

Pressure transmitters **PCE-28, PCE-28P**, differential pressure transmitters **PRE-28** in marine version comply with regulations of Det Norske Veritas to be used for ships, vessels and oil rigs and are marked with the following certificate marking:

**DNV No. A-11308**

and class of installation standards: temperature: class **C**, humidity : class **B**, vibrations: class **B**, EMC class: **B**, housing : class **C**.

Certificate was issued by Det Norske Veritas, Veritasveien 1, 1322 Hovik, Norway

Warsaw, 15th October 2009

Piotr Pawelec  
*P. Pawelec*  
Quality Manager